



**West Midlands**  
Interchange

**Four Ashes Ltd**

**Calibration Certificates for Position N6 – June/July 2018 Survey**



## Documentation Métrologique Metrological documentation

**FUSION 11403**

Date d'émission : **12/06/2017**  
Date of issue :

Référence Document : NOT1536  
Nom : Documentation métrologique - *Metrological documentation* FRGB

**[www.acoemgroup.com](http://www.acoemgroup.com)**  
**[support@acoemgroup.com](mailto:support@acoemgroup.com)**

## TABLE DES MATIERES TABLE OF CONTENT

Chapitre 1.	Constat de vérification <i>Verification certificate</i> .....	5
Chapitre 2.	Certificat d'étalonnage <i>Calibration certificate</i> .....	11
Chapitre 3.	Certificat de conformité <i>Conformity certificate</i> .....	21



# Chapitre 1.

## CONSTAT DE VERIFICATION

### VERIFICATION CERTIFICATE

CV-DTE-L-17-PVE-49602

DELIVRE PAR :  
ISSUED BY :

ACOEM  
Service Métrologie

69760 LIMONEST  
France

INSTRUMENT VERIFIE  
INSTRUMENT CHECKED

Désignation :  
Designation :

**Sonomètre Intégrateur-Moyenueur**  
**Integrating-Averaging Sound Level Meter**

Constructeur :  
Manufacturer :

**01dB**

Type :  
Type :

**FUSION**


N° de serie :  
Serial number : **11403**

N° d'identification :  
Identification number

Date d'émission :  
Date of issue : **12/06/2017**

Ce constat comprend **5** pages  
This certificate includes **5** pages

LE RESPONSABLE METROLOGIQUE  
DU LABORATOIRE  
HEAD OF THE METROLOGY LAB  
François MAGAND



LA REPRODUCTION DE CE CONSTAT N'EST AUTORISEE  
QUE SOUS LA FORME DE FAC-SIMILE PHOTOGRAPHIQUE INTEGRAL

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED OTHER  
THAN IN FULL BY PHOTOGRAPHIC PROCESS

CE DOCUMENT NE PEUT PAS ETRE UTILISE EN LIEU  
ET PLACE D'UN CERTIFICAT D'ETALONNAGE. CE DOCUMENT  
EST REALISE SUIVANT LES RECOMMANDATIONS DU  
FASCICULE DE DOCUMENTATION X 07-011.

THIS DOCUMENT CAN'T BE USED AS CALIBRATION  
CERTIFICATE. IT IS COMPLIANT WITH THE X 07-011 STANDARD  
RECOMMENDATIONS.



**IDENTIFICATION :**

IDENTIFICATION:

	Sonomètre <i>Sound level meter</i>	Préamplificateur <i>Preamplifier</i>	Microphone <i>Microphone</i>
Constructeur : <i>Manufacturer</i>	01dB		GRAS
Type : <i>Type</i>	FUSION	Interne - Internal	40CE
Numéro de série : <i>Serial number</i>	11403		259481

**PROGRAMME DE VERIFICATION :**

VERIFICATION PROGRAM:

Ce sonomètre a été vérifié sur les caractéristiques suivantes:

- Réponse en fréquence du sonomètre
- Linéarité
- Pondérations fréquentielles A-B-C-Z
- Bruit de fond
- Filtre 1/1 et 1/3 octave

*This sound level meter has been verified on its following characteristics:*

- *Frequency response of the sound level meter*
- *Linearity*
- *A-B-C-Z Weighting*
- *Background noise*
- *1/1 and 1/3 Octave filter*

**METHODE DE VERIFICATION :**

VERIFICATION METHOD:

L'appareil est vérifié dans une salle climatisée. Les caractéristiques sont vérifiées étalonnées avec un multimètre et un générateur étalonnés en amplitude et en fréquence. Des corrections constructeurs sont appliquées pour prendre en compte les effets des accessoires et du boîtier selon la norme IEC 61672-3

*The instrument is controlled in an air conditioned room. The other characteristics are verified with multimeter and generator calibrated in amplitude and in frequency. Some manufacturer's corrections have been applied to account the acoustical effect from the case of the sound level meter and his accessories (IEC 61672-3).*

**CONDITIONS DE VERIFICATION :**

VERIFICATION CONDITIONS:

Date de l'étalonnage : .12 - 6 - 2017.

*Date of Calibration (french format)*

Nom de l'opérateur : **Didier Lafond**

*Operator Name*

Instruction d'étalonnage : **P118-NOT-01**

*Calibration instruction*

Pression atmosphérique : **98,63 kPa**

*Static pressure*

Température : **24,4 °C**

*Temperature*

Taux d'humidité relative : **43,3 %HR**

*Relative humidity*



**MOYENS DE MESURE UTILISES POUR LA VERIFICATION :**

INSTRUMENTS USED FOR VERIFICATION:

Désignation	Constructeur	Type	N° de série	N° d'identification
Designation	Manufacturer	Type	Serial number	Identification number
Générateur de fonction / Waveform generator	Helwet-Packard	HP 33120 A	US36028745	APM 1163
Boite à décades / Decade box	01dB-Metravib	OUT1694	1605202	APM 5541

Tous les moyens de mesure utilisés sont raccordés aux étalons de référence de la société ACOEM. Les étalons de référence de la société ACOEM sont raccordés aux étalons nationaux par un étalonnage COFRAC. La liste de ces étalons est disponible sur simple demande auprès du responsable métrologique du laboratoire.

*All the measuring instruments are calibrated using the ACOEM reference standards. ACOEM reference standards are calibrated with COFRAC certificate of calibration. The reference standard list is available on simple request to the head of the Metrology Lab.*

**RESULTATS :**

RESULTS:

Le jugement de conformité de chaque test est établi suivant les tolérances données dans les normes suivantes :

*Conformity decision has been taken with the tolerance descriptions in the following standards:*

IEC 61260

IEC 61672-1 classe

1

ANSI S1.11 class

ANSI S1.4 class

1

Linéarité  
Linearity

Description <i>Description</i>	Résultat <i>Result</i>
Linéarité <i>Linearity</i>	Conforme <i>Compliant</i>

Pondérations fréquentielles A-B-C-Z  
A-B-C-Z Weightings

Description <i>Description</i>	Résultat <i>Result</i>
Pondération fréquentielle <i>Frequency weighting</i>	Conforme <i>Compliant</i>

Bruit de fond  
Background noise

Description <i>Description</i>	Résultat <i>Result</i>
Bruit de fond <i>Noise level</i>	Conforme <i>Compliant</i>



Filtre d'octave  
1/1 Octave filter

Description <i>Description</i>	Résultat <i>Result</i>
Fréquence centrale filtre 1/1 octave <i>1/1 Octave filter central frequency attenuation</i>	Conforme <i>Compliant</i>

Filtre de 1/3 d'octave  
1/3 Octave filter

Description <i>Description</i>	Résultat <i>Result</i>
Fréquence centrale filtre 1/3 octave <i>1/3 Octave filter central frequency attenuation</i>	Conforme <i>Compliant</i>

Les données liées au DMK01 sont issues de la réponse en fréquence du microphone associé à l'influence typique du DMK01.

The DMK01's results describes the association of the microphone acoustical response with the typical DMK01 influence.

Fin du constat de vérification End of verification certificate



# Chapitre 2.

## CERTIFICAT D'ETALONNAGE

### CALIBRATION CERTIFICATE

CE-DTE-L-17-PVE-49602

DELIVRE PAR :  
ISSUED BY :

ACOEM  
Service Métrologie

69760 LIMONEST  
France

INSTRUMENT ETALONNE  
CALIBRATED INSTRUMENT

Désignation :  
Designation :

**Sonomètre Intégrateur-Moyenneur**  
**Integrating-Averaging Sound Level Meter**

Constructeur :  
Manufacturer :

**01dB**

Type :  
Type :

**FUSION**

N° de serie :  
Serial number : **11403**

N° d'identification :  
Identification number

Date d'émission :  
Date of issue : **12/06/2017**

Ce certificat comprend **10** Pages  
This certificate includes **10** Pages

LE RESPONSABLE METROLOGIQUE  
DU LABORATOIRE  
HEAD OF THE METROLOGY LAB  
François MAGAND

LA REPRODUCTION DE CE CERTIFICAT N'EST AUTORISEE QUE  
SOUS LA FORME DE FAC-SIMILE PHOTOGRAPHIQUE INTEGRAL.  
THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL  
BY PHOTOGRAPHIC PROCESS

CE CERTIFICAT EST CONFORME AU FASCICULE DE  
DOCUMENTATION FD X 07-012.  
THIS CERTIFICATE IS COMPLIANT WITH THE FD X 07-012  
STANDARD DOCUMENTATION



**IDENTIFICATION :**

IDENTIFICATION:

	Sonomètre <i>Sound level meter</i>	Préamplificateur <i>Preamplifier</i>	Microphone <i>Microphone</i>
Constructeur : <i>Manufacturer</i>	01dB		GRAS
Type : <i>Type</i>	FUSION	Interne - Internal	40CE
Numéro de série : <i>Serial number</i>	11403		259481

**PROGRAMME D'ETALONNAGE :**

CALIBRATION PROGRAM:

Ce Sonomètre a été étalonné sur les caractéristiques suivantes :

- Réponse en fréquence du sonomètre en champ libre
- Linéarité
- Pondérations fréquentielles A-B-C-Z

*The Sound level meter has been calibrated on the following characteristics:*

- *Free field frequency response of the sound level meter*
- *Linearity*
- *A-B-C-Z frequency weightings*

**METHODE D'ETALONNAGE :**

CALIBRATION METHOD:

L'appareil est étalonné dans une salle climatisée. Les caractéristiques sont étalonnées avec un multimètre et un générateur étalonnés en amplitude et en fréquence. Des corrections constructeurs sont appliquées pour prendre en compte les effets des accessoires et du boîtier selon la norme IEC 61672-3

*The instrument is calibrated in an air conditioned room.. The other characteristics are verified with multimeter and generator calibrated in amplitude and in frequency. Some manufacturer's corrections have been applied to account the acoustical effect from the case of the sound level meter and his accessories (IEC 61672-3).*

**CONDITIONS D'ETALONNAGE :**

CALIBRATION CONDITIONS:

Date de l'étalonnage : .12 - 6 - 2017.

*Date of Calibration (french format)*

Nom de l'opérateur : **Didier Lafond**

*Operator Name*

Instruction d'étalonnage : **P118-NOT-01**

*Calibration instruction*

Pression atmosphérique : **98,63 kPa**

*Static pressure*

Température : **24,4 °C**

*Temperature*

Taux d'humidité relative : **43,3 %HR**

*Relative humidity*



**MOYENS DE MESURES UTILISES POUR L'ETALONNAGE :**

INSTRUMENTS USED FOR CALIBRATION:

Désignation	Constructeur	Type	N° de série	N° d'identification
Designation	Manufacturer	Type	Serial number	Identification number
Générateur de fonction / Waveform generator	Helwet-Packard	HP 33120 A	US36028745	APM 1163
Boite à décades / Decade box	01dB-Metravib	OUT1694	1605202	APM 5541

Tous les moyens de mesure utilisés sont raccordés aux étalons de référence de la société ACOEM. Les étalons de référence de la société ACOEM sont raccordés aux étalons nationaux par un étalonnage COFRAC. La liste de ces étalons est disponible sur simple demande auprès du responsable métrologique du laboratoire.

*All the measuring instruments are calibrated using the ACOEM reference standards. ACOEM reference standards are calibrated to national standard with COFRAC certificate of calibration. The reference standards list is available on simple request to the head of the Metrology lab.*

**RESULTATS :**

RESULTS:

Les incertitudes élargies mentionnées sont celles correspondant à deux incertitudes types ( $k=2$ ). Les incertitudes types sont calculées en tenant compte des différentes composantes d'incertitudes, étalons de référence, moyens d'étalonnage, conditions d'environnement, contribution de l'instrument étalonné, répétabilité ...

*Mentioned expanded uncertainties correspond to two standard uncertainty types ( $k=2$ ). Standard uncertainties are calculated including different uncertainty components, reference standards, instruments used, environmental conditions, calibrated instrument contribution, repeatability...*



**Pondération fréquentielle**  
**Frequency Weighting**

<b>Pondération fréquentielle (voie interne) - Frequency weighting (primary)</b>					
0° Short windscreen	<b>Z</b>	<b>A</b>	<b>B</b>	<b>C</b>	Incertitude uncertainty (dB)
63 Hz	-0,7	-27,0	-10,1	-1,6	0,45
125 Hz	-0,6	-16,9	-4,9	-0,8	0,45
250 Hz	-0,6	-9,3	-2,0	-0,6	0,29
500 Hz	-0,4	-3,6	-0,6	-0,3	0,29
1000 Hz	-0,3	-0,3	-0,3	-0,3	0,29
2000 Hz	0,4	1,6	0,3	0,2	0,29
4000 Hz	-0,5	0,4	-1,3	-1,4	0,39
8000 Hz	-1,5	-3,1	-4,8	-5,0	0,61
16000 Hz	-1,6	-13,6	-15,4	-15,5	0,61

**Linéarité**  
Linearity

Linéarité (voie principale) <i>Linearity (Primary channel)</i>	Valeur nominale <i>Nominal value</i> ( dB )	Valeur affichée <i>Displayed value</i> ( dB )	Incertitudes <i>Uncertainty</i> ( dB )
Leq 35 dBZ / 8000 Hz	35,0	34,9	0,23
Leq 40 dBZ / 8000 Hz	40,0	40,1	0,23
Leq 50 dBZ / 8000 Hz	50,0	50,0	0,20
Leq 60 dBZ / 8000 Hz	60,0	60,0	0,20
Leq 70 dBZ / 8000 Hz	70,0	70,0	0,20
Leq 80 dBZ / 8000 Hz	80,0	80,0	0,20
Leq 90 dBZ / 8000 Hz	90,0	90,0	0,20
Leq 100 dBZ / 8000 Hz	100,0	100,0	0,20
Leq 110 dBZ / 8000 Hz	110,0	109,9	0,20
Leq 120 dBZ / 8000 Hz	120,0	119,8	0,20
Leq 130 dBZ / 8000 Hz	130,0	129,8	0,20
Leq 134 dBZ / 8000 Hz	134,0	133,7	0,20
Leq 134 dBA / 8000 Hz	134,0	133,7	0,20
Leq 130 dBA / 8000 Hz	130,0	129,7	0,20
Leq 120 dBA / 8000 Hz	120,0	119,8	0,20
Leq 110 dBA / 8000 Hz	110,0	109,8	0,20
Leq 100 dBA / 8000 Hz	100,0	100,0	0,20
Leq 90 dBA / 8000 Hz	90,0	90,0	0,20
Leq 80 dBA / 8000 Hz	80,0	80,0	0,20
Leq 70 dBA / 8000 Hz	70,0	70,0	0,20
Leq 60 dBA / 8000 Hz	60,0	60,0	0,20
Leq 50 dBA / 8000 Hz	50,0	50,0	0,20
Leq 40 dBA / 8000 Hz	40,0	40,0	0,23
Leq 30 dBA / 8000 Hz	30,0	30,0	0,23
Leq 26 dBA / 8000 Hz	26,0	26,1	0,23




**Filtre**  
 Filter

Filtre par bande d'octave (Voie principale) <i>Octave filter (primary channel)</i>	Valeur nominale <i>Nominal value</i> ( dB )	Valeur affichée <i>Displayed value</i> ( dB )	Incertitudes <i>Uncertainty</i> ( dB )
Leq 110 dB / 1/1 Octave / 31,5 Hz	110,0	109,9	0,5
Leq 110 dB / 1/1 Octave / 63 Hz	110,0	109,9	0,5
Leq 110 dB / 1/1 Octave / 125 Hz	110,0	109,9	0,5
Leq 110 dB / 1/1 Octave / 250 Hz	110,0	110,0	0,3
Leq 110 dB / 1/1 Octave / 500 Hz	110,0	110,0	0,3
Leq 110 dB / 1/1 Octave / 1000 Hz	110,0	110,0	0,3
Leq 110 dB / 1/1 Octave / 2000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/1 Octave / 4000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/1 Octave / 8000 Hz	110,0	109,9	0,4

Filtre tiers d'octave (Voie principale) <i>Third octave filter (Primary channel)</i>	Valeur nominale <i>Nominal value</i> ( dB )	Valeur affichée <i>Displayed value</i> ( dB )	Incertitudes <i>Uncertainty</i> ( dB )
Leq 110 dB / 1/3 Octave / 25 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 31,5 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 40 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 50 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 63 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 80 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 100 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 125 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 160 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 200 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 250 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 315 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 400 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 500 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 630 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 800 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 1000 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 1250 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 1600 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 2000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 2500 Hz	110,0	110,1	0,4
Leq 110 dB / 1/3 Octave / 3150 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 4000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 5000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 6300 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 8000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 10000 Hz	110,0	109,9	0,6



**Réponse acoustique**  
Acoustic response



Frequency (Hz)	Channel Z (dB)	Channel A (dB)	Channel B (dB)	Channel C (dB)
63	-1.0	-1.0	-1.0	-1.0
125	-1.0	-1.0	-1.0	-1.0
250	-1.0	-1.0	-1.0	-1.0
500	-1.0	-1.0	-1.0	-1.0
1000	-1.0	-1.0	-1.0	-1.0
2000	-1.0	-1.0	-1.0	-1.0
4000	-1.0	-1.0	-1.0	-1.0
8000	-1.0	-1.0	-1.0	-1.0
16000	-1.0	-1.0	-1.0	-1.0

Frequency (Hz)	Channel Z (dB)	Channel A (dB)	Channel B (dB)	Channel C (dB)
63	-1.0	-1.0	-1.0	-1.0
125	-1.0	-1.0	-1.0	-1.0
250	-1.0	-1.0	-1.0	-1.0
500	-1.0	-1.0	-1.0	-1.0
1000	-1.0	-1.0	-1.0	-1.0
2000	-1.0	-1.0	-1.0	-1.0
4000	-1.0	-1.0	-1.0	-1.0
8000	-1.0	-1.0	-1.0	-1.0
16000	-1.0	-1.0	-1.0	-1.0



### OPTION DMK 01 (1/3)

Les données liées au DMK01 sont issues de la réponse en fréquence du microphone associé à l'influence typique du DMK01.

The DMK01's results describes the association of the microphone acoustical response with the typical DMK01 influence.


Filtre par bande d'octave (DMK 01) <i>Octave filter (with DMK01)</i>	Valeur nominale <i>Nominal value</i> ( dB )	Valeur affichée <i>Displayed value</i> ( dB )	Incertitudes <i>Uncertainty</i> ( dB )
Leq 110 dB / 1/1 Octave / 31,5 Hz	110,0	109,9	0,5
Leq 110 dB / 1/1 Octave / 63 Hz	110,0	109,9	0,5
Leq 110 dB / 1/1 Octave / 125 Hz	110,0	109,9	0,5
Leq 110 dB / 1/1 Octave / 250 Hz	110,0	110,0	0,3
Leq 110 dB / 1/1 Octave / 500 Hz	110,0	110,0	0,3
Leq 110 dB / 1/1 Octave / 1000 Hz	110,0	110,0	0,3
Leq 110 dB / 1/1 Octave / 2000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/1 Octave / 4000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/1 Octave / 8000 Hz	110,0	109,9	0,4

Filtre tiers d'octave (DMK 01) <i>Third octave filter (with DMK01)</i>	Valeur nominale <i>Nominal value</i> ( dB )	Valeur affichée <i>Displayed value</i> ( dB )	Incertitudes <i>Uncertainty</i> ( dB )
Leq 110 dB / 1/3 Octave / 25 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 31,5 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 40 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 50 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 63 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 80 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 100 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 125 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 160 Hz	110,0	110,0	0,5
Leq 110 dB / 1/3 Octave / 200 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 250 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 315 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 400 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 500 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 630 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 800 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 1000 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 1250 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 1600 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 2000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 2500 Hz	110,0	110,1	0,4
Leq 110 dB / 1/3 Octave / 3150 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 4000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 5000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 6300 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 8000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 10000 Hz	110,0	109,9	0,6



**OPTION DMK 01 (2/3)**

Linéarité (avec DMK01) <i>Linearity (with DMK01)</i>	Valeur nominale <i>Nominal value</i> ( dB )	Valeur affichée <i>Displayed value</i> ( dB )	Incertitudes <i>Uncertainty</i> ( dB )
Leq 35 dBZ / 8000 Hz	35,0	35,0	0,23
Leq 40 dBZ / 8000 Hz	40,0	40,1	0,23
Leq 50 dBZ / 8000 Hz	50,0	50,0	0,20
Leq 60 dBZ / 8000 Hz	60,0	60,0	0,20
Leq 70 dBZ / 8000 Hz	70,0	70,0	0,20
Leq 80 dBZ / 8000 Hz	80,0	80,0	0,20
Leq 90 dBZ / 8000 Hz	90,0	90,0	0,20
Leq 100 dBZ / 8000 Hz	100,0	100,0	0,20
Leq 110 dBZ / 8000 Hz	110,0	109,8	0,20
Leq 120 dBZ / 8000 Hz	120,0	119,7	0,20
Leq 130 dBZ / 8000 Hz	130,0	129,7	0,20
Leq 134 dBZ / 8000 Hz	134,0	133,7	0,20
Leq 134 dBA / 8000 Hz	134,0	133,7	0,20
Leq 130 dBA / 8000 Hz	130,0	129,8	0,20
Leq 120 dBA / 8000 Hz	120,0	119,7	0,20
Leq 110 dBA / 8000 Hz	110,0	109,9	0,20
Leq 100 dBA / 8000 Hz	100,0	100,0	0,20
Leq 90 dBA / 8000 Hz	90,0	90,0	0,20
Leq 80 dBA / 8000 Hz	80,0	80,0	0,20
Leq 70 dBA / 8000 Hz	70,0	70,0	0,20
Leq 60 dBA / 8000 Hz	60,0	60,0	0,20
Leq 50 dBA / 8000 Hz	50,0	50,0	0,20
Leq 40 dBA / 8000 Hz	40,0	39,9	0,23
Leq 30 dBA / 8000 Hz	30,0	30,1	0,23
Leq 26 dBA / 8000 Hz	26,0	26,1	0,23





### OPTION DMK 01 (3/3)

Pondération fréquentielle (avec DMK01) Frequency weighting (with DMK01)			
Z	0° RA0208 + Short windscreen	90° RA208 + short windscreen	Incertitude uncertainty
63 Hz	-0,3	-0,1	0,45
125 Hz	-0,2	-0,1	0,45
250 Hz	-0,1	0,0	0,29
500 Hz	-0,1	-0,1	0,29
1000 Hz	0,0	0,0	0,29
2000 Hz	0,2	-0,1	0,29
4000 Hz	0,0	-0,1	0,39
8000 Hz	-1,1	-1,1	0,61
16000 Hz	-3,5	-8,5	0,61
A	0° RA0208 + Short windscreen	90° RA208 + short windscreen	Incertitude uncertainty
63 Hz	-26,6	-26,4	0,45
125 Hz	-16,4	-16,3	0,45
250 Hz	-8,8	-8,7	0,29
500 Hz	-3,3	-3,4	0,29
1000 Hz	0,0	0,0	0,29
2000 Hz	1,4	1,1	0,29
4000 Hz	0,9	0,8	0,39
8000 Hz	-2,7	-2,7	0,61
16000 Hz	-15,5	-20,4	0,61
B	0° RA0208 + Short windscreen	90° RA208 + short windscreen	Incertitude uncertainty
63 Hz	-9,6	-9,5	0,45
125 Hz	-4,5	-4,3	0,45
250 Hz	-1,5	-1,4	0,29
500 Hz	-0,4	-0,4	0,29
1000 Hz	0,0	0,0	0,29
2000 Hz	0,1	-0,2	0,29
4000 Hz	-0,8	-0,9	0,39
8000 Hz	-4,5	-4,5	0,61
16000 Hz	-17,3	-22,3	0,61
C	0° RA0208 + Short windscreen	90° RA208 + short windscreen	Incertitude uncertainty
63 Hz	-1,1	-1,0	0,45
125 Hz	-0,4	-0,2	0,45
250 Hz	-0,1	0,0	0,29
500 Hz	-0,1	-0,1	0,29
1000 Hz	0,0	0,0	0,29
2000 Hz	0,1	-0,3	0,29
4000 Hz	-0,9	-1,0	0,39
8000 Hz	-4,6	-4,6	0,61
16000 Hz	-17,4	-22,4	0,61

Fin du certificat d'étalonnage End of calibration certificate



# Chapitre 3.

## CERTIFICAT DE CONFORMITE

### CONFORMITY CERTIFICATE

---

CC-DTE-L-17-PVE-49602

Nous, fabricant  
We, manufacturer

**Acoem**  
200, Chemin des Ormeaux  
F 69578 LIMONEST Cedex- FRANCE

déclarons sous notre seule responsabilité que le produit suivant :  
declare under our own responsibility that the following equipment:

Désignation : **Sonomètre Intégrateur Moyenneur**  
Designation: **Integrating-Averaging Sound level meter**

Référence : **FUSION**  
Reference:

Numéro de série : **11403**  
Serial Number:

est conforme aux dispositions des normes suivantes :  
complies with the requirements of the following standards:

	Norme Standard	Classe Class	Edition du Edition of
<b>Sonomètre :</b>	IEC 60651	1	10-2000
<b>Sound level meter :</b>	IEC 60804	1	10-2000
	IEC 61672-1	1	09-2013
	IEC 61260	1	07-1995-2011
	ANSI S1.11	1	2004
	ANSI S1.4	1	1983-1985

et répond en tout point, après vérification et essais, aux exigences spécifiées, aux normes et règlements applicables, sauf exceptions, réserves ou dérogations énumérées dans la présente déclaration de conformité.

*After testing and verification, this device satisfies all specified requirements and applicable standards and regulations apart from exceptions, reservations, or exemptions listed in this conformance certificate.*

Date LE REFERENT METROLOGIE ACOUSTIQUE  
Date PAR DELEGATION  
THE REFERENT ACOUSTIC METROLOGY  
Bertrand LEROY

**12/06/2017**





**Calibration Certificates for Position N7 – June/July 2018 Survey**

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 1 of 3

## APPROVED SIGNATORIES

Claire Lomax [x]      Andy Moorhouse [ ]  
Gary Phillips [ ]      Danny McCaul [ ]



0801

## acoustic calibration laboratory

The University of Salford, Salford, Greater Manchester, M5 4WT, UK  
<http://www.acoustics.salford.ac.uk>  
t 0161 295 3030/0161 295 3319 f 0161 295 4456 e c.lomax1@salford.ac.uk

University of  
**Salford**  
MANCHESTER

Certificate Number: 03556/1

Date of Issue: 18 December 2017

### PERIODIC TEST OF A SOUND LEVEL METER to IEC 61672-3:2006

FOR:	Acoustic 1 The Barns Overdale Manordeilo Llandeilo Carmarthenshire, SA19 7BD
FOR THE ATTENTION OF:	Steve Thomas
PERIODIC TEST DATE:	18/12/2017
TEST PROCEDURE:	CTP12 (Laboratory Manual)

#### Sound Level Meter Details

Manufacturer	01dB	
Model	CUBE	
Serial number	11118	
Class	1	
Hardware version	LIS001B	Application FW: 2.40 Metrology FW: 2.12

Associated Items	Microphone	Preamplifier
Manu	GRAS	01dB
Model	40CD	PRE22
Serial Number	260866	1610458

Test Engineer (initial):

*GP*

Name:

Gary Phillips

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*



# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 2 of 3

Certificate Number: 03556/1

Date of Issue: 18 December 2017

Procedures from IEC 61672-3: 2006 and TPS 49 Edition 2 June 2009 were used to perform the periodic tests.

The manufacturer's instruction manual was marked as follows: DOC1144 June 2014 G - CUBE User Manual GB. Adjustment data used to adjust the sound levels indicated in response to the application of a multi-frequency sound calibrator to sound levels equivalent to those that would be indicated in response to plane, progressive sound waves were obtained from the manufacturer. The sound level meter calibration check frequency is 1000 Hz, the reference sound pressure level is 94 dB. As this instrument only has a single range, this range is the reference level range.

The environmental conditions in the laboratory at the start of the test were:

Static pressure 102.453 kPa  $\pm$  0.022 kPa; air temperature 21.8 °C  $\pm$  0.3 °C; relative humidity 39.7 %  $\pm$  1.8 %.

The initial response of the instrument to application of the suitable laboratory sound calibrator was 93.9 dB (C). The instrument was then adjusted to indicate 93.8 dB (C). This indication was obtained from the calibration certificate of the calibrator and information in the manufacturer's instruction manual specified in this certificate, when the instrument is configured for use with the external input, using the DMK01 weatherproof outdoor microphone unit, supplied RAL135-10M microphone extension cable and the following instrument settings; Microphone input: External, Microphone type: 40CD, Reference direction: 90°, High-pass filter: 10 Hz, Nose cone: Yes.

With the microphone replaced by an electrical input device with a similar capacitance to that of the electrical input device specified in the manufacturer's instruction manual, the levels of self-generated noise were:

**A: 10.8 dB\***  
**B: 10.2 dB\***  
**C: 11.5 dB\***  
**Z: 16.8 dB\***

\* Under-range indicated on instrument display.

The environmental conditions in the laboratory at the end of the test were:

Static pressure 102.566 kPa  $\pm$  0.022 kPa; air temperature 21.9 °C  $\pm$  0.3 °C; relative humidity 42.3 %  $\pm$  1.8 %.

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organization responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 3 of 3

Certificate Number: 03556/1

Date of Issue: 18 December 2017

As the actual frequency response of the microphone was unavailable, the typical frequency response for the model of microphone has been used to correct the level differences determined in the electrical signal test of frequency weighting.

The instrument failed to meet the requirements for the test of electrical signal tests of frequency weightings at 250 Hz, 500 Hz and 1 kHz for the A, B, C and Z-weightings, as the uncertainty of measurement exceeded the maximum permitted value due to a significant contribution from data supplied by the manufacturer. If the manufacturer's uncertainty data were not included, the meter would meet the requirements of the Standard.

Instruments used in the verification procedure were traceable to *National Standards*. The multi-frequency calibrator method was employed in the acoustical tests of a frequency weighting.

*The uncertainty evaluation has been carried out in accordance with UKAS requirements. All measurement results are retained at the acoustic calibration laboratory for at least four years.*

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*



**Calibration Certificates for Position N8 – June/July 2018 Survey**



## Documentation Métrologique Metrological documentation

**CUBE 11165**

Date d'émission : **16/05/2017**  
Date of issue :

Référence Document : NOT1536  
Nom : Documentation métrologique - *Metrological documentation* FRGB

**[www.acoemgroup.com](http://www.acoemgroup.com)**  
**[support@acoemgroup.com](mailto:support@acoemgroup.com)**



## TABLE DES MATIERES TABLE OF CONTENT

Chapitre 1.	Constat de verification <i>Verification certificate</i> .....	5
Chapitre 2.	Certificat d'étalonnage <i>Calibration certificate</i> .....	9
Chapitre 3.	Certificat de conformité <i>Conformity certificate</i> .....	15

# Chapitre 1.

## CONSTAT DE VERIFICATION

### VERIFICATION CERTIFICATE

---

CV-DTE-L-17-PVE-49149

DELIVRE PAR :  
ISSUED BY :

ACOEM  
Service Métrologie

69760 LIMONEST  
France

INSTRUMENT VERIFIE  
INSTRUMENT CHECKED

Désignation :  
Designation :

**Sonomètre Intégrateur-Moyenneur**  
**Integrating-Averaging Sound Level Meter**

Constructeur :  
Manufacturer :

**01dB**

Type :  
Type :

**CUBE**

N° de serie :  
Serial number :

**11165**

N° d'identification :  
Identification number

Date d'émission :  
Date of issue :

**16/05/2017**

Ce constat comprend  
This certificate includes

**5** pages  
pages

LE RESPONSABLE METROLOGIQUE  
DU LABORATOIRE  
HEAD OF THE METROLOGY LAB



François MAGAND

LA REPRODUCTION DE CE CONSTAT N'EST AUTORISEE  
QUE SOUS LA FORME DE FAC-SIMILE PHOTOGRAPHIQUE INTEGRAL

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED OTHER  
THAN IN FULL BY PHOTOGRAPHIC PROCESS

CE DOCUMENT NE PEUT PAS ETRE UTILISE EN LIEU  
ET PLACE D'UN CERTIFICAT D'ETALONNAGE. CE DOCUMENT  
EST REALISE SUIVANT LES RECOMMANDATIONS DU  
FASCICULE DE DOCUMENTATION X 07-011.

THIS DOCUMENT CAN'T BE USED AS CALIBRATION  
CERTIFICATE. IT IS COMPLIANT WITH THE X 07-011 STANDARD  
RECOMMENDATIONS.



**IDENTIFICATION :***IDENTIFICATION:*

	Sonomètre <i>Sound level meter</i>	Préamplificateur <i>Preamplifier</i>	Microphone <i>Microphone</i>
Constructeur : <i>Manufacturer</i>	01dB	01dB	GRAS
Type : <i>Type</i>	CUBE	PRE22	40CD
Numéro de série : <i>Serial number</i>	11165	161406	287995

**PROGRAMME DE VERIFICATION :***VERIFICATION PROGRAM:*

Ce sonomètre a été vérifié sur les caractéristiques suivantes:

- Réponse en fréquence du sonomètre
- Linéarité
- Pondérations fréquentielles A-B-C-Z
- Bruit de fond
- Filtre 1/1 et 1/3 octave

*This sound level meter has been verified on its following characteristics:*

- *Frequency response of the sound level meter*
- *Linearity*
- *A-B-C-Z Weighting*
- *Background noise*
- *1/1 and 1/3 Octave filter*

**METHODE DE VERIFICATION :***VERIFICATION METHOD:*

L'appareil est vérifié dans une salle climatisée. Les caractéristiques sont vérifiées étalonnées avec un multimètre et un générateur étalonnés en amplitude et en fréquence. Des corrections constructeurs sont appliquées pour prendre en compte les effets des accessoires et du boîtier selon la norme IEC 61672-3

*The instrument is controlled in an air conditioned room. The other characteristics are verified with multimeter and generator calibrated in amplitude and in frequency. Some manufacturer's corrections have been applied to account the acoustical effect from the case of the sound level meter and his accessories (IEC 61672-3).*

**CONDITIONS DE VERIFICATION :***VERIFICATION CONDITIONS:*

Date de l'étalonnage : 16/05/2017  
*Date of Calibration (french format)*

Nom de l'opérateur : **Didier Lafond**  
*Operator Name*

Instruction d'étalonnage : **P118-NOT-01**  
*Calibration instruction*

Pression atmosphérique : **99,37 kPa**  
*Static pressure*

Température : **24,9 °C**  
*Temperature*

Taux d'humidité relative : **37 %HR**  
*Relative humidity*

**MOYENS DE MESURE UTILISES POUR LA VERIFICATION :***INSTRUMENTS USED FOR VERIFICATION:*

Désignation	Constructeur	Type	N° de série	N° d'identification
Designation	Manufacturer	Type	Serial number	Identification number
Générateur de fonction / Waveform generator	Helwet-Packard	HP 33120 A	US36028745	APM 1163
Boite à décades / Decade box	01dB-Metravib	OUT1694	1605202	APM 5541

Tous les moyens de mesure utilisés sont raccordés aux étalons de référence de la société Acoem. Les étalons de référence de la société Acoem sont raccordés aux étalons nationaux par un étalonnage COFRAC. La liste de ces étalons est disponible sur simple demande auprès du responsable métrologique du laboratoire.

*All the measuring instruments are calibrated using the Acoem reference standards. Acoem reference standards are calibrated with COFRAC certificate of calibration. The reference standard list is available on simple request to the head of the Metrology Lab.*

**RESULTATS :***RESULTS:*

Le jugement de conformité de chaque test IEC 61260  
est établi suivant les tolérances données IEC 61672-1 classe 1  
dans les normes suivantes :

*Conformity decision has been taken with the* ANSI S1.11 class 1  
*tolerance descriptions in the following*  
*standards:* ANSI S1.4 class



Linéarité  
*Linearity*

Description <i>Description</i>	Résultat <i>Result</i>
Linéarité <i>Linearity</i>	Conforme <i>Compliant</i>

Pondérations fréquentielles A-B-C-Z  
*A-B-C-Z Weightings*

Description <i>Description</i>	Résultat <i>Result</i>
Pondération fréquentielle <i>Frequency weighting</i>	Conforme <i>Compliant</i>

Bruit de fond  
*Background noise*

Description <i>Description</i>	Résultat <i>Result</i>
Bruit de fond <i>Noise level</i>	Conforme <i>Compliant</i>

Filtre d'octave  
*1/1 Octave filter*

Description <i>Description</i>	Résultat <i>Result</i>
Fréquence centrale filtre 1/1 octave <i>1/1 Octave filter central frequency attenuation</i>	Conforme <i>Compliant</i>

Filtre de 1/3 d'octave  
*1/3 Octave filter*

Description <i>Description</i>	Résultat <i>Result</i>
Fréquence centrale filtre 1/3 octave <i>1/3 Octave filter central frequency attenuation</i>	Conforme <i>Compliant</i>

Fin du constat de vérification    End of verification certificate

# Chapitre 2.

## CERTIFICAT D'ETALONNAGE

### CALIBRATION CERTIFICATE

---

CE-DTE-L-17-PVE-49149

DELIVRE PAR :  
ISSUED BY :

ACOEM  
Service Métrologie

69760 LIMONEST  
France

INSTRUMENT ETALONNE  
CALIBRATED INSTRUMENT

Désignation :  
Designation :

**Sonomètre Intégrateur-Moyenneur**  
**Integrating-Averaging Sound Level Meter**

Constructeur :  
Manufacturer :

**01dB**

Type :  
Type :

**CUBE**

N° de serie :  
Serial number :

**11165**


N° d'identification :  
Identification number

Date d'émission :  
Date of issue :

**16/05/2017**

Ce certificat comprend **6** Pages  
This certificate includes **6** Pages

LE RESPONSABLE METROLOGIQUE  
DU LABORATOIRE  
HEAD OF THE METROLOGY LAB



François MAGAND

LA REPRODUCTION DE CE CERTIFICAT N'EST AUTORISEE QUE  
SOUS LA FORME DE FAC-SIMILE PHOTOGRAPHIQUE INTEGRAL.  
THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL  
BY PHOTOGRAPHIC PROCESS

CE CERTIFICAT EST CONFORME AU FASCICULE DE  
DOCUMENTATION FD X 07-012.  
THIS CERTIFICATE IS COMPLIANT WITH THE FD X 07-012  
STANDARD DOCUMENTATION



**IDENTIFICATION :**

IDENTIFICATION:

	Sonomètre <i>Sound level meter</i>	Préamplificateur <i>Preamplifier</i>	Microphone <i>Microphone</i>
Constructeur : <i>Manufacturer</i>	01dB	01dB	GRAS
Type : <i>Type</i>	CUBE	PRE22	40CD
Numéro de série : <i>Serial number</i>	11165	161406	287995

**PROGRAMME D'ETALONNAGE :**

CALIBRATION PROGRAM:

Ce Sonomètre a été étalonné sur les caractéristiques suivantes :

- Réponse en fréquence du sonomètre en champ libre
- Linéarité
- Pondérations fréquentielles A-B-C-Z

*The Sound level meter has been calibrated on the following characteristics:*

- *Free field frequency response of the sound level meter*
- *Linearity*
- *A-B-C-Z frequency weightings*

**METHODE D'ETALONNAGE :**

CALIBRATION METHOD:

L'appareil est étalonné dans une salle climatisée. Les caractéristiques sont étalonnées avec un multimètre et un générateur étalonnés en amplitude et en fréquence. Des corrections constructeurs sont appliquées pour prendre en compte les effets des accessoires et du boîtier selon la norme IEC 61672-3

*The instrument is calibrated in an air conditioned room.. The other characteristics are verified with multimeter and generator calibrated in amplitude and in frequency. Some manufacturer's corrections have been applied to account the acoustical effect from the case of the sound level meter and his accessories (IEC 61672-3).*

**CONDITIONS D'ETALONNAGE :**

CALIBRATION CONDITIONS:

Date de l'étalonnage : 16/05/2017  
*Date of Calibration* (french format)  
 Nom de l'opérateur : **Didier Lafond**  
*Operator Name*  
 Instruction d'étalonnage : **P118-NOT-01**  
*Calibration instruction*

Pression atmosphérique : **99,37 kPa**  
*Static pressure*

Température : **24,9 °C**  
*Temperature*

Taux d'humidité relative : **37 %HR**  
*Relative humidity*

**MOYENS DE MESURES UTILISES POUR L'ETALONNAGE :***INSTRUMENTS USED FOR CALIBRATION:*

Désignation	Constructeur	Type	N° de série	N° d'identification
Designation	Manufacturer	Type	Serial number	Identification number
Générateur de fonction / Waveform generator	Helwet-Packard	HP 33120 A	US36028745	APM 1163
Boite à décades / Decade box	01dB-Metravib	OUT1694	1605202	APM 5541

Tous les moyens de mesure utilisés sont raccordés aux étalons de référence de la société Acoem . Les étalons de référence de la société Acoem sont raccordés aux étalons nationaux par un étalonnage COFRAC. La liste de ces étalons est disponible sur simple demande auprès du responsable métrologique du laboratoire.

*All the measuring instruments are calibrated using the Acoem reference standards. Acoem reference standards are calibrated to national standard with COFRAC certificate of calibration. The reference standards list is available on simple request to the head of the Metrology lab.*

**RESULTATS :***RESULTS:*

Les incertitudes élargies mentionnées sont celles correspondant à deux incertitudes types ( $k=2$ ). Les incertitudes types sont calculées en tenant compte des différentes composantes d'incertitudes, étalons de référence, moyens d'étalonnage, conditions d'environnement, contribution de l'instrument étalonné, répétabilité ...

*Mentioned expanded uncertainties correspond to two standard uncertainty types ( $k=2$ ). Standard uncertainties are calculated including different uncertainty components, reference standards, instruments used, environmental conditions, calibrated instrument contribution, repeatability...*





**Pondération fréquentielle**

*Frequency Weighting*

DMK 0°: Pondération fréquentielle ; Frequency weighting					
Fréquence Frequency (Hz)	Z	A	B	C	Incertitude uncertainty (dB)
63 Hz	-0,2	-26,5	-9,6	-1,1	0,45
125 Hz	-0,1	-16,3	-4,4	-0,3	0,45
250 Hz	-0,1	-8,7	-1,5	-0,1	0,29
500 Hz	-0,2	-3,4	-0,4	-0,1	0,29
1000 Hz	0,0	0,0	0,0	0,0	0,29
2000 Hz	0,0	1,2	-0,1	-0,2	0,29
4000 Hz	-0,4	0,5	-1,2	-1,3	0,39
8000 Hz	-1,4	-3,0	-4,8	-4,9	0,61
16000 Hz	-7,0	-19,0	-20,8	-20,9	0,61

DMK 90°: Pondération fréquentielle ; Frequency weighting					
Fréquence Frequency (Hz)	Z	A	B	C	Incertitude uncertainty (dB)
63 Hz	-0,3	-26,5	-9,7	-1,1	0,45
125 Hz	-0,2	-16,4	-4,4	-0,4	0,45
250 Hz	-0,2	-8,8	-1,5	-0,2	0,29
500 Hz	-0,2	-3,5	-0,5	-0,2	0,29
1000 Hz	0,0	0,0	0,0	0,0	0,29
2000 Hz	0,0	1,2	-0,1	-0,2	0,29
4000 Hz	-0,5	0,5	-1,2	-1,3	0,39
8000 Hz	-0,7	-2,3	-4,1	-4,2	0,61
16000 Hz	-5,4	-17,4	-19,2	-19,3	0,61



**Linéarité***Linearity*

Linéarité (voie principale) <i>Linearity (Primary channel)</i>	Valeur nominale <i>Nominal value</i> ( dB )	Valeur affichée <i>Displayed value</i> ( dB )	Incertitudes <i>Uncertainty</i> ( dB )
Leq 35 dBZ / 8000 Hz	35,0	35,1	0,23
Leq 45 dBZ / 8000 Hz	45,0	45,0	0,23
Leq 50 dBZ / 8000 Hz	50,0	50,0	0,20
Leq 70 dBZ / 8000 Hz	70,0	70,0	0,20
Leq 90 dBZ / 8000 Hz	90,0	90,0	0,20
Leq 110 dBZ / 8000 Hz	110,0	109,9	0,20
Leq 130 dBZ / 8000 Hz	130,0	129,7	0,20
Leq 134 dBZ / 8000 Hz	134,0	133,8	0,20
Leq 134 dBA / 8000 Hz	134,0	133,8	0,20
Leq 130 dBA / 8000 Hz	130,0	129,7	0,20
Leq 110 dBA / 8000 Hz	110,0	109,9	0,20
Leq 90 dBA / 8000 Hz	90,0	90,0	0,20
Leq 70 dBA / 8000 Hz	70,0	70,0	0,20
Leq 50 dBA / 8000 Hz	50,0	50,0	0,20
Leq 30 dBA / 8000 Hz	30,0	30,0	0,20
Leq 26 dBA / 8000 Hz	26,0	26,1	0,20



**Filtre**

Filter

Filtre par bande d'octave (Voie principale) <i>Octave filter (primary channel)</i>	Valeur nominale <i>Nominal value</i> ( dB )	Valeur affichée <i>Displayed value</i> ( dB )	Incertitudes <i>Uncertainty</i> ( dB )
Leq 110 dB / 1/1 Octave / 31,5 Hz	110,0	109,9	0,5
Leq 110 dB / 1/1 Octave / 63 Hz	110,0	109,9	0,5
Leq 110 dB / 1/1 Octave / 125 Hz	110,0	110,0	0,5
Leq 110 dB / 1/1 Octave / 250 Hz	110,0	110,0	0,3
Leq 110 dB / 1/1 Octave / 500 Hz	110,0	110,0	0,3
Leq 110 dB / 1/1 Octave / 1000 Hz	110,0	110,0	0,3
Leq 110 dB / 1/1 Octave / 2000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/1 Octave / 4000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/1 Octave / 8000 Hz	110,0	110,0	0,4

Filtre tiers d'octave (Voie principale) <i>Third octave filter (Primary channel)</i>	Valeur nominale <i>Nominal value</i> ( dB )	Valeur affichée <i>Displayed value</i> ( dB )	Incertitudes <i>Uncertainty</i> ( dB )
Leq 110 dB / 1/3 Octave / 25 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 31,5 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 40 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 50 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 63 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 80 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 100 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 125 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 160 Hz	110,0	109,9	0,5
Leq 110 dB / 1/3 Octave / 200 Hz	110,0	109,9	0,3
Leq 110 dB / 1/3 Octave / 250 Hz	110,0	109,9	0,3
Leq 110 dB / 1/3 Octave / 315 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 400 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 500 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 630 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 800 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 1000 Hz	110,0	110,0	0,3
Leq 110 dB / 1/3 Octave / 1250 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 1600 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 2000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 2500 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 3150 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 4000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 5000 Hz	110,0	110,0	0,4
Leq 110 dB / 1/3 Octave / 6300 Hz	110,0	109,9	0,4
Leq 110 dB / 1/3 Octave / 8000 Hz	110,0	109,9	0,4
Leq 110 dB / 1/3 Octave / 10000 Hz	110,0	109,9	0,6

Fin du certificat d'étalonnage End of calibration certificate

# Chapitre 3.

## CERTIFICAT DE CONFORMITE

### CONFORMITY CERTIFICATE

---

CC-DTE-L-17-PVE-49149

Nous, fabricant  
*We, manufacturer*

**Acoem**  
200, Chemin des Ormeaux  
F 69578 LIMONEST Cedex- FRANCE

déclarons sous notre seule responsabilité que le produit suivant :  
*declare under our own responsibility that the following equipment:*

Désignation : **Sonomètre Intégrateur Moyenneur**  
*Designation: Integrating-Averaging Sound level meter*

Référence : **CUBE**  
*Reference:*

Numéro de série : **11165**  
*Serial Number:*

est conforme aux dispositions des normes suivantes :  
*complies with the requirements of the following standards:*


	Norme <i>Standard</i>	Classe <i>Class</i>	Edition du <i>Edition of</i>
<b>Sonomètre :</b>	IEC 60651	1	10-2000
<b>Sound level meter :</b>	IEC 60804	1	10-2000
	IEC 61672-1	1	09-2013
	IEC 61260	1	07-1995-2011
	ANSI S1.11	1	2004
	ANSI S1.4	1	1983-1985

et répond en tout point, après vérification et essais, aux exigences spécifiées, aux normes et règlements applicables, sauf exceptions, réserves ou dérogations énumérées dans la présente déclaration de conformité.

*After testing and verification, this device satisfies all specified requirements and applicable standards and regulations apart from exceptions, reservations, or exemptions listed in this conformance certificate.*

Date LE REFERENT METROLOGIE ACOUSTIQUE  
PAR DELEGATION  
Date THE REFERENT ACOUSTIC METROLOGY  
Bertrand LEROY

**16/05/2017**



**Calibration Certificates for Position N9 – June/July 2018 Survey**



# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 1 of 3

## APPROVED SIGNATORIES

Claire Lomax [x]      Andy Moorhouse [ ]  
Gary Phillips [ ]      Danny McCaul [ ]



0801

## acoustic calibration laboratory

The University of Salford, Salford, Greater Manchester, M5 4WT, UK  
<http://www.acoustics.salford.ac.uk>  
t 0161 295 3030/0161 295 3319 f 0161 295 4456 e c.lomax1@salford.ac.uk

University of  
**Salford**  
MANCHESTER

Certificate Number: 03556/3

Date of Issue: 19 December 2017

### PERIODIC TEST OF A SOUND LEVEL METER to IEC 61672-3:2006

FOR:	Acoustic 1 The Barns Overdale Manordeilo Llandeilo Carmarthenshire, SA19 7BD
FOR THE ATTENTION OF:	Steve Thomas
PERIODIC TEST DATE:	19/12/2017
TEST PROCEDURE:	CTP12 (Laboratory Manual)

#### Sound Level Meter Details

Manufacturer	01dB	
Model	CUBE	
Serial number	11110	
Class	1	
Hardware version	LIS001B	Application FW: 2.40. Metrology FW: 2.12

Associated Items	Microphone	Preamplifier
Manu	GRAS	01dB
Model	40CD	PRE22
Serial Number	260893	1610541

Test Engineer (initial):

*GP*

Name:

Gary Phillips

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 2 of 3

Certificate Number: 03556/3

Date of Issue: 19 December 2017

Procedures from IEC 61672-3: 2006 and TPS 49 Edition 2 June 2009 were used to perform the periodic tests.

The manufacturer's instruction manual was marked as follows: DOC1144 June 2014 G - CUBE User Manual GB. Adjustment data used to adjust the sound levels indicated in response to the application of a multi-frequency sound calibrator to sound levels equivalent to those that would be indicated in response to plane, progressive sound waves were obtained from the manufacturer. The sound level meter calibration check frequency is 1000 Hz, the reference sound pressure level is 94 dB. As this instrument only has a single range, this range is the reference level range.

The environmental conditions in the laboratory at the start of the test were:

Static pressure 102.840 kPa  $\pm$  0.022 kPa; air temperature 22.0 °C  $\pm$  0.3 °C; relative humidity 36.1 %  $\pm$  1.8 %.

The initial response of the instrument to application of the suitable laboratory sound calibrator was 93.8 dB (C). No adjustment of the instrument was required. This indication was obtained from the calibration certificate of the calibrator and information in the manufacturer's instruction manual specified in this certificate, when the instrument is configured for use with the external input, using the DMK01 weatherproof outdoor microphone unit, supplied RAL135-5M microphone extension cable and the following instrument settings; Microphone input: External, Microphone type: 40CD, Reference direction: 90°, High-pass filter: 10 Hz, Nose cone: Yes.

With the microphone replaced by an electrical input device with a similar capacitance to that of the electrical input device specified in the manufacturer's instruction manual, the levels of self-generated noise were:

**A: 10.7 dB\***  
**B: 10.0 dB\***  
**C: 11.1 dB\***  
**Z: 16.9 dB\***

\* Under-range indicated on instrument display.

The environmental conditions in the laboratory at the end of the test were: Static pressure 102.902 kPa  $\pm$  0.022 kPa; air temperature 22.5 °C  $\pm$  0.3 °C; relative humidity 36.4 %  $\pm$  1.8 %.

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organization responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 3 of 3

Certificate Number: 03556/3

Date of Issue: 19 December 2017

As the actual frequency response of the microphone was unavailable, the typical frequency response for the model of microphone has been used to correct the level differences determined in the electrical signal test of frequency weighting.

The instrument failed to meet the requirements for the test of electrical signal tests of frequency weightings at 250 Hz, 500 Hz and 1 kHz for the A, B, C and Z-weightings, as the uncertainty of measurement exceeded the maximum permitted value due to a significant contribution from data supplied by the manufacturer. If the manufacturer's uncertainty data were not included, the meter would meet the requirements of the Standard.

Instruments used in the verification procedure were traceable to *National Standards*. The multi-frequency calibrator method was employed in the acoustical tests of a frequency weighting.

*The uncertainty evaluation has been carried out in accordance with UKAS requirements. All measurement results are retained at the acoustic calibration laboratory for at least four years.*

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*



**Calibration Certificates for Position N10 – June/July 2018 Survey**

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 1 of 3

## APPROVED SIGNATORIES

Claire Lomax [x]      Andy Moorhouse [ ]  
Gary Phillips [ ]      Danny McCaul [ ]



0801

## acoustic calibration laboratory

The University of Salford, Salford, Greater Manchester, M5 4WT, UK  
<http://www.acoustics.salford.ac.uk>  
t 0161 295 3030/0161 295 3319 f 0161 295 4456 e c.lomax1@salford.ac.uk

University of  
**Salford**  
MANCHESTER

Certificate Number: 03666/2

Date of Issue: 15 March 2018

### PERIODIC TEST OF A SOUND LEVEL METER to IEC 61672-3:2006

FOR:	Acoustic 1 The Barns Overdale Manordeilo Llandeilo Carmarthenshire SA19 7BD
FOR THE ATTENTION OF:	Steve Thomas
PERIODIC TEST DATE:	13/03/2018
TEST PROCEDURE:	CTP12 (Laboratory Manual)

#### Sound Level Meter Details

Manufacturer	01dB	
Model	DUO	
Serial number	10522	
Class	1	
Hardware version	3F2D3D	Application FW: 2.38. Metrology FW: 2.12

Associated Items	Microphone	Preamplifier
Manu	GRAS	01dB
Model	40CD	PRE22
Serial Number	154578	10329

Test Engineer (initial):

*GP*

Name:

Gary Phillips

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 2 of 3

Certificate Number: 03666/2

Date of Issue: 15 March 2018

Procedures from IEC 61672-3: 2006 and TPS 49 Edition 2 June 2009 were used to perform the periodic tests. The manufacturer's instruction manual was marked as follows: DOC1112 June 2014 G - DUO User Manual GB. Adjustment data used to adjust the sound levels indicated in response to the application of a multi-frequency sound calibrator to sound levels equivalent to those that would be indicated in response to plane, progressive sound waves were obtained from the manufacturer. The sound level meter calibration check frequency is 1000 Hz; the reference sound pressure level is 94 dB. As this instrument only has a single range, this range is the reference level range.

The environmental conditions in the laboratory at the start of the test were:

Static pressure 100.264 kPa  $\pm$  0.022 kPa; air temperature 22.5 °C  $\pm$  0.3 °C; relative humidity 38.6 %  $\pm$  1.8 %.

The initial response of the instrument to application of the associated sound calibrator was 93.4 dB (C). The instrument was then adjusted to indicate 93.8 dB (C). This indication was obtained from the calibration certificate of the calibrator and information in the manufacturer's instruction manual specified in this certificate, when the instrument is configured for use with the external input, using the DMK01 weatherproof outdoor microphone unit, supplied RAL135-5M microphone extension cable and the following instrument settings; Microphone input: External, Microphone type: 40CD, Reference direction: 90°, High-pass filter: 10 Hz, Nose cone: Yes.

With the microphone replaced by an electrical input device with a similar capacitance to that of the electrical input device specified by the manufacturer, the levels of self-generated noise were:

**A: 12.8 dB\***  
**B: 12.2 dB\***  
**C: 13.4 dB\***  
**Z: 18.9 dB\***

\* Under-range indicated on instrument display

The environmental conditions in the laboratory at the end of the test were:

Static pressure 100.271 kPa  $\pm$  0.022 kPa; air temperature 23.1 °C  $\pm$  0.3 °C; relative humidity 34.7 %  $\pm$  1.8 %.

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organization responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*



# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 3 of 3

Certificate Number: 03666/2

Date of Issue: 15 March 2018

The instrument failed to meet the requirements for the test of electrical signal tests of frequency weightings at 250 Hz, 500 Hz and 1 kHz for the A, B, C and Z-weightings, as the uncertainty of measurement exceeded the maximum permitted value due to a significant contribution from data supplied by the manufacturer. If the manufacturer's uncertainty data were not included, the meter would meet the requirements of the Standard.

As the actual frequency response of the microphone was unavailable, the typical frequency response for the model of microphone has been used to correct the level differences determined in the electrical signal test of frequency weighting.

Instruments used in the verification procedure were traceable to *National Standards*. The multi-frequency calibrator method was employed in the acoustical tests of a frequency weighting.

*The uncertainty evaluation has been carried out in accordance with UKAS requirements. All measurement results are retained at the acoustic calibration laboratory for at least four years.*

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

**Calibration Certificates for Position N11 – June/July 2018 Survey**

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 1 of 3

## APPROVED SIGNATORIES

Claire Lomax [x]      Andy Moorhouse [ ]  
Gary Phillips [ ]      Danny McCaul [ ]

## acoustic calibration laboratory

The University of Salford, Salford, Greater Manchester, M5 4WT, UK  
<http://www.acoustics.salford.ac.uk>  
t 0161 295 3030/0161 295 3319 f 0161 295 4456 e c.lomax1@salford.ac.uk



UKAS  
CALIBRATION

0801

University of  
**Salford**  
MANCHESTER

Certificate Number: 03224/1

Date of Issue: 3 May 2017

## PERIODIC TEST OF A SOUND LEVEL METER to IEC 61672-3:2006

FOR:	Acoustic 1 The Barns Overdale Manordeilo Llandeilo Carmarthenshire SA19 7BD
FOR THE ATTENTION OF:	Steve Thomas
PERIODIC TEST DATE:	03/05/2017
TEST PROCEDURE:	CTP12 (Laboratory Manual)

### Sound Level Meter Details

Manufacturer	01dB	
Model	CUBE	
Serial number	10694	
Class	1	
Hardware version	LIS001B	Application FW: 2.34. Metrology FW: 2.10

Associated Items	Microphone	Preamplifier
Manu	GRAS	01dB
Model	40CD	PRE22
Serial Number	224223	11118

Test Engineer (initial):

*GP*

Name:

Gary Phillips

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 2 of 3

Certificate Number: 03224/1

Date of Issue: 3 May 2017

Procedures from IEC 61672-3: 2006 and TPS 49 Edition 2 June 2009 were used to perform the periodic tests. The manufacturer's instruction manual was marked as follows: DOC1144 June 2014 G - CUBE User Manual GB. Adjustment data used to adjust the sound levels indicated in response to the application of a multi-frequency sound calibrator to sound levels equivalent to those that would be indicated in response to plane, progressive sound waves were obtained from the manufacturer. The sound level meter calibration check frequency is 1000 Hz, the reference sound pressure level is 94 dB. As this instrument only has a single range, this range is the reference level range.

The environmental conditions in the laboratory at the start of the test were:

Static pressure 102.124 kPa  $\pm$  0.017 kPa, air temperature 21.5 °C  $\pm$  0.3 °C, relative humidity 42.4 %  $\pm$  1.9%.

The initial response of the instrument to application of the suitable laboratory sound calibrator was 93.8 dB (C). The instrument was then adjusted to indicate 93.7 dB (C). This indication was obtained from the calibration certificate of the calibrator and information in the manufacturer's instruction manual specified in this certificate, when the instrument is configured for use with the external input, using the DMK01 weatherproof outdoor microphone unit, supplied RAL135-10m microphone extension cable and the following instrument settings; Microphone input: External, Microphone type: 40CD, Reference direction: 90°, High-pass filter: 10 Hz, Nose cone: Yes.

With the microphone replaced by an electrical input device with a similar capacitance to that of the electrical input device specified in the manufacturer's instruction manual, the levels of self-generated noise were:

**A: 11.2 dB\***  
**B: 10.8 dB\***  
**C: 12.1 dB\***  
**Z: 17.3 dB\***

\* Under-range indicated on instrument display.

The environmental conditions in the laboratory at the end of the test were:

Static pressure 102.056 kPa  $\pm$  0.017 kPa, air temperature 24.0 °C  $\pm$  0.3 °C, relative humidity 38.1 %  $\pm$  1.9%.

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organization responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.



# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 3 of 3

Certificate Number: 03224/1

Date of Issue: 3 May 2017

The instrument failed to meet the requirements for the test of electrical signal tests of frequency weightings at 250 Hz, 500 Hz and 1 kHz for the A, B, C and Z-weightings, as the uncertainty of measurement exceeded the maximum permitted value due to a significant contribution from data supplied by the manufacturer. If the manufacturer's uncertainty data were not included, the meter would meet the requirements of the Standard.

The microphone corrections applied as specified in 12.6 of IEC 61672-3:2006 were obtained from a frequency response measured by this Laboratory using the electrostatic actuator method. This response in isolation is not covered by our UKAS accreditation.

As the actual frequency response of the microphone was unavailable, the typical frequency response for the model of microphone has been used to correct the level differences determined in the electrical signal test of frequency weighting.

Instruments used in the verification procedure were traceable to *National Standards*. The multi-frequency calibrator method was employed in the acoustical tests of a frequency weighting.

*The uncertainty evaluation has been carried out in accordance with UKAS requirements. All measurement results are retained at the acoustic calibration laboratory for at least four years.*

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

**Calibration Certificates for Position N12 – June/July 2018 Survey**

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 1 of 3

## APPROVED SIGNATORIES

Claire Lomax [x]      Andy Moorhouse [ ]  
Gary Phillips [ ]      Danny McCaul [ ]

## acoustic calibration laboratory

The University of Salford, Salford, Greater Manchester, M5 4WT, UK  
<http://www.acoustics.salford.ac.uk>  
t 0161 295 3030/0161 295 3319 f 0161 295 4456 e c.lomax1@salford.ac.uk



0801

University of  
**Salford**  
MANCHESTER

Certificate Number: 03112/6

Date of Issue: 27 March 2017

## PERIODIC TEST OF A SOUND LEVEL METER to IEC 61672-3:2006

FOR:	Acoustic 1 The Barns Overdale Manordeilo Llandeilo Carmarthenshire SA19 7BD
FOR THE ATTENTION OF:	Steve Thomas
PERIODIC TEST DATE:	24 <sup>th</sup> and 27 <sup>th</sup> March 2017
TEST PROCEDURE:	CTP12 (Laboratory Manual)

### Sound Level Meter Details

Manufacturer	01dB	
Model	CUBE	
Serial number	10619	
Class	1	
Hardware version	LIS001A	Application FW: 2.35. Metrology FW: 2.12

Associated Items	Microphone	Preamplifier
Manu	GRAS	01dB
Model	40CD	PRE22
Serial Number	207246	10730

Test Engineer (initial):

*GP*

Name:

Gary Phillips

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 2 of 3

Certificate Number: 03112/6

Date of Issue: 27 March 2017

Procedures from IEC 61672-3: 2006 and TPS 49 Edition 2 June 2009 were used to perform the periodic tests.

The manufacturer's instruction manual was marked as follows: DOC1144 June 2014 G - CUBE User Manual GB. Adjustment data used to adjust the sound levels indicated in response to the application of a multi-frequency sound calibrator to sound levels equivalent to those that would be indicated in response to plane, progressive sound waves were obtained from the manufacturer. The sound level meter calibration check frequency is 1000 Hz, the reference sound pressure level is 94 dB. As this instrument only has a single range, this range is the reference level range.

The environmental conditions in the laboratory at the start of the test were:

Static pressure 102.712 kPa  $\pm$  0.017 kPa, air temperature 22.3 °C  $\pm$  0.3 °C, relative humidity 37.6 %  $\pm$  1.9%.

The initial response of the instrument to application of the suitable laboratory sound calibrator was 93.9 dB (C). The instrument was then adjusted to indicate 93.7 dB (C). This indication was obtained from the calibration certificate of the calibrator and information in the manufacturer's instruction manual specified in this certificate, when the instrument is configured for use with the external input, using the DMK01 weatherproof outdoor microphone unit, supplied RAL135-10m microphone extension cable and the following instrument settings; Microphone input: External, Microphone type: 40CD, Reference direction: 90°, High-pass filter: 10 Hz, Nose cone: Yes.

With the microphone replaced by an electrical input device with a similar capacitance to that of the electrical input device specified in the manufacturer's instruction manual, the levels of self-generated noise were:

**A: 11.6 dB\***  
**B: 11.0 dB\***  
**C: 12.3 dB\***  
**Z: 17.7 dB\***

\* Under-range indicated on instrument display.

The environmental conditions in the laboratory at the end of the test were:

Static pressure 101.920 kPa  $\pm$  0.017 kPa, air temperature 22.7 °C  $\pm$  0.3 °C, relative humidity 38.6 %  $\pm$  1.9%.

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organization responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*



# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 3 of 3

Certificate Number: 03112/6

Date of Issue: 27 March 2017

As the actual frequency response of the microphone was unavailable, the typical frequency response for the model of microphone has been used to correct the level differences determined in the electrical signal test of frequency weighting.

The instrument failed to meet the requirements for the test of electrical signal tests of frequency weightings at 250 Hz, 500 Hz and 1 kHz for the A, B, C and Z-weightings, as the uncertainty of measurement exceeded the maximum permitted value due to a significant contribution from data supplied by the manufacturer. If the manufacturer's uncertainty data were not included, the meter would meet the requirements of the Standard.

Instruments used in the verification procedure were traceable to *National Standards*. The multi-frequency calibrator method was employed in the acoustical tests of a frequency weighting.

*The uncertainty evaluation has been carried out in accordance with UKAS requirements. All measurement results are retained at the acoustic calibration laboratory for at least four years.*

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

**Calibration Certificate for Calibrator used at N1, N2, N4, N6, N7, N8 and N12 –  
June/July 2018 Survey**

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 1 of 2

## APPROVED SIGNATORIES

Claire Lomax [ ]      Andy Moorhouse [ ]

Gary Phillips [x]      Danny McCaul [ ]

## acoustic calibration laboratory

The University of Salford, Salford, Greater Manchester, M5 4WT, UK  
<http://www.acoustics.salford.ac.uk>  
t 0161 295 3030/0161 295 3319 f 0161 295 4456 e c.lomax1@salford.ac.uk



0801

University of  
**Salford**  
MANCHESTER

Certificate Number: 03725/1

Date of Issue: 20 April 2018

## CALIBRATION OF A SOUND CALIBRATOR

FOR:                    Acoustic 1  
                              The Barns  
                              Overdale  
                              Manordeilo  
                              Llandeilo  
                              Carmarthenshire  
                              SA19 7BD

FOR THE ATTENTION OF:      Steve Thomas

DESCRIPTION:                    Calibrator with housing for one-inch  
   microphones and adaptor type BAC21 for  
   half-inch microphones.

MANUFACTURER:                01 dB

TYPE:                                CAL21

SERIAL NUMBER:                35054818 (2015)

DATE OF CALIBRATION:        17/04/2018

TEST PROCEDURE:                CTP06 (Laboratory Manual)

Test Engineer (initial): CL

Name: Claire Lomax

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 2 of 2

Certificate Number: 03725/1

Date of Issue: 20 April 2018

## MEASUREMENTS

The sound pressure level generated by the calibrator was measured using a calibrated, WS2P condenser microphone specified in this certificate. The calibration was carried out with the calibrator in the half-inch configuration.

Five determinations of the sound pressure level, frequency and total distortion were made.

The manufacturer states that automatic compensation is applied for the effects of changes in atmospheric pressure.

Consult the manufacturer's instruction manual for any corrections, due to slight differences in microphone load volumes, to the stated level which may be required when using this sound calibrator with other microphone models.

## RESULTS

Coupler configuration:	Half-inch
Microphone type:	GRAS 40AG
Output level (dB re 20 $\mu$ Pa):	94.16 dB $\pm$ 0.11 dB
Frequency (Hz):	1001.40 Hz $\pm$ 0.01 %
Total Distortion (%):	1.57 % $\pm$ 0.39 %

Average environmental conditions at the time of measurement were:

Pressure:	100.966 kPa $\pm$ 0.023 kPa
Temperature:	22.4 °C $\pm$ 0.4 °C
Relative humidity:	43.5 % $\pm$ 1.8 %

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.*

*All measurement results are retained at the acoustic calibration laboratory for at least four years.*



**Calibration Certificate for Calibrator used at N8, N9, N10, N11 – June/July 2018  
Survey**

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 1 of 2

## APPROVED SIGNATORIES

Claire Lomax [ ]      Andy Moorhouse [ ]

Gary Phillips [x]      Danny McCaul [ ]



0801

## acoustic calibration laboratory

The University of Salford, Salford, Greater Manchester, M5 4WT, UK  
<http://www.acoustics.salford.ac.uk>  
t 0161 295 3030/0161 295 3319 f 0161 295 4456 e c.lomax1@salford.ac.uk

University of  
**Salford**  
MANCHESTER

Certificate Number: 03759/1

Date of Issue: 3 May 2018

## CALIBRATION OF A SOUND CALIBRATOR

FOR:                      Resound Acoustics Limited  
                                 13 Rother Street  
                                 Stratford-upon-Avon  
                                 Warwickshire  
                                 CV37 6LU

FOR THE ATTENTION OF:      Mike Brownstone

DESCRIPTION:                      Calibrator with housing for one-inch  
                                 microphones and adaptor type BAC21 for  
                                 half-inch microphones.

MANUFACTURER:                      01 dB

TYPE:                                      CAL21

SERIAL NUMBER:                      34134139 (2013)

DATE OF CALIBRATION:              01/05/2018

TEST PROCEDURE:                      CTP06 (Laboratory Manual)

Test Engineer (initial): CL

Name: Claire Lomax

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 2 of 2

Certificate Number: 03759/1

Date of Issue: 3 May 2018

## MEASUREMENTS

The sound pressure level generated by the calibrator was measured using a calibrated WS2P condenser microphone as specified in this certificate. The calibration was carried out with the calibrator in the half-inch configuration.

Five determinations of the sound pressure level, frequency and total distortion were made.

The manufacturer states that automatic compensation is applied for the effects of changes in atmospheric pressure.

## RESULTS

Coupler configuration:	Half-inch
Microphone type:	GRAS 40AG
Output level (dB re 20 $\mu$ Pa):	94.03 dB $\pm$ 0.11 dB
Frequency (Hz):	1002.60 Hz $\pm$ 0.01 %
Total Distortion (%):	1.60 % $\pm$ 0.39 %

Average environmental conditions at the time of measurement were:

Pressure:	99.903 kPa $\pm$ 0.023 kPa
Temperature:	22.3 $^{\circ}$ C $\pm$ 0.4 $^{\circ}$ C
Relative humidity:	37.8 % $\pm$ 1.8 %

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.*

*All measurement results are retained at the acoustic calibration laboratory for at least four years.*

Figure A13A.3.56: Full survey results, August 2016, Position N1






Figure A13A.3.57: Full survey results, August 2016, Position N2






Figure A13A.3.59: Full survey results, August 2016, Position N5




Figure A13A.3.60: Full survey results, October 2016, Position N6




Figure A13A.3.61: Full survey results, August 2016, Position N7






Figure A13A.3.62: Full survey results, August 2016, Position N8




Figure A13A.3.63: Full survey results, August 2016, Position N9




Figure A13A.3.64: Full survey results, January 2017, Position N1




Figure A13A.3.65: Full survey results, January 2017, Position N2




Figure A13A.3.66: Full survey results, January 2017, Position N4






Figure A13A.3.67: Full survey results, January 2017, Position N5




Figure A13A.3.68: Full survey results, January 2017, Position N6




Figure A13A.3.69: Full survey results, January 2017, Position N7




Figure A13A.3.70: Full survey results, January 2017, Position N8





Figure A13A.3.71: Full survey results, January 2017, Position N9





**Calibration Certificates for Position N1 – August 2016 Survey**

# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

DATE OF ISSUE: 2 June 2016

CERTIFICATE NUMBER: 160922

BS EN ISO  
9001:2008  
APPROVED  
BY  
**LRQA**

CERT No 953910



Home Farm Industrial Park  
Norwich Road  
Marsham  
Norfolk  
NR10 5PQ  
Tel: +44 1603 279557  
Fax: +44 1603 278008

Page 1 of 8  
Approved Signatory

Electronically Authorised Document

- P K CLARK  
 R J WADE  
 M A FROST  
 M S PARDOE

<b><u>Customer</u></b>	TECL LIMITED O/B OF SPL ACOUSTICS LTD
<b><u>Order No</u></b>	TECL/P00064/010616
<b><u>Equipment Description</u></b>	SOUND LEVEL METER
<b><u>Manufacturer</u></b>	LARSON DAVIS
<b><u>Model</u></b>	820
<b><u>Serial No</u></b>	1350
<b><u>Ident No</u></b>	SPL A
<b><u>Date Of Calibration</u></b>	2 JUNE 2016

## **INSTRUMENT CONDITION**

<b><u>Adjustments Made</u></b>	YES
<b><u>Repairs Made</u></b>	NO

## **ENVIRONMENT**

The instrument was placed in the laboratory environment for a minimum period of 4 hours and was operated prior to calibration.

Measurements were made in ambient conditions of 22°C ± 3°C and 45% ± 15% RH.

## **PROCEDURE**

Measurements were performed in accordance with the in house laboratory procedure 0265. All equipment used has been calibrated/verified against measurement standards or reference equipment traceable to International or National Measurement Standards as specified in our control procedure WI64.

The results attached to this certificate refer to measurements made at the time of test and not to the instrument's ability to maintain calibration.

The attached results are a true record of the levels required to return the instrument to the original stated manufacturer's specification and accuracy where known.

# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER

160922

Page 2 of 8

## INSTRUMENTS USED

### EQUIPMENT

Bruel & Kjaer 4226

### SERIAL No

2952856

### CERTIFICATE No

U19912

### CAL DUE

26 Oct 2016

## Notes:

### MEASUREMENT UNCERTAINTIES

The expanded uncertainty quoted refers to the measured values only, with no account being taken of the instruments ability to maintain its calibration. The expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%.

PARAMETER	RANGE	EXPANDED UNCERTAINTY
d.c. Resistance	0.01 $\Omega$ - 400M $\Omega$	$\pm$ 409ppm
	400M $\Omega$ - 1T $\Omega$	$\pm$ 1%
d.c. Voltage	0V - 1kV	$\pm$ 79ppm
d.c. Voltage	1.01kV-15kV	$\pm$ 2.2%
d.c. Current	0mA - 20A	$\pm$ 437ppm
a.c. Voltage	0mV - 1.05kV	$\pm$ 1.2%
a.c. Current	0mA - 20A	$\pm$ 0.5%
Frequency	0.5Hz - 20GHz	$\pm$ 0.1ppm
Capacitance	0.5nF - 40mF	$\pm$ 1.1%
Time	0 - 1 Hour	$\pm$ 1s
Distortion	10mV - 100V	$\pm$ 1.4mV
Temperature (Dry Block)	-30 $^{\circ}$ C - 350 $^{\circ}$ C	$\pm$ 1%
Temperature (Simulation)	-270 $^{\circ}$ C - 1800 $^{\circ}$ C	$\pm$ 0.57%
Pressure	10mBar - 35Bar	$\pm$ 0.04%
Torque	0.1 - 1100Nm	$\pm$ 0.5%
Weight	2g - 157kg	$\pm$ 0.03%
Humidity	0% - 90%	$\pm$ 1%
Shock & Impulse Hammers		$\pm$ 4%
Spring Hammers		$\pm$ 0.015J
Sound	Frequency	$\pm$ 0.06%
Sound	Level	$\pm$ 0.16dB
Tachometers	60rpm - 96000rpm	$\pm$ 0.1%
Anemometers	2.5m/s to 15m/s	$\pm$ 2.0%
Vibration Meters	10Hz - 1kHz	$\pm$ 5%
Vibration Calibrators		$\pm$ 3%
Mechanical Measurement	<200mm	$\pm$ 0.0003mm
	>200mm	$\pm$ 0.002mm
Inductance		$\pm$ 0.1%
Power (VA)		$\pm$ 1%
Power (RF)		$\pm$ 0.5dB

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER

**160922**


## RESULT SHEET 0265 - 820 SOUND LEVEL METER

### AS FOUND

#### 1) FREQUENCY RESPONSE

##### A Weighted

Applied Freq	Applied Level	A Weighting Typical	Indicated Pressure Response	Free Field Response	Units	Error % Pressure	Error % FFR
31.5Hz	104.0	64.471	64.0	64.0	dBA	-0.73	-0.73
63Hz	104.0	77.777	77.4	77.4	dBA	-0.48	-0.48
125Hz	104.0	87.811	87.7	87.7	dBA	-0.13	-0.13
250Hz	104.0	95.325	95.2	95.2	dBA	-0.13	-0.13
500Hz	104.0	100.752	100.5	100.5	dBA	-0.25	-0.25
1kHz	104.0	104.000	103.5	103.7	dBA	-0.48	-0.31
2kHz	104.0	105.202	104.4	104.9	dBA	-0.76	-0.33
4kHz	104.0	104.964	103.5	104.6	dBA	-1.39	-0.35
8kHz	104.0	102.855	99.8	103.1	dBA	-2.97	0.21
12.5kHz	104.0	99.750	93.2	99.6	dBA	-6.57	-0.13
16kHz	104.0	97.300	89.2	97.2	dBA	-8.32	-0.08





# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910




CERTIFICATE NUMBER

**160922**

## RESULT SHEET 0265 - 820 SOUND LEVEL METER

### 1) FREQUENCY RESPONSE (Continued) C Weighted

Applied Freq	Applied Level	C Weighting Typical	Indicated Pressure Response	Free Field Response	Units	Error % Pressure	Error % FFR
31.5Hz	104.0	100.969	100.0	100.0	dB(A)	-0.96	-0.96
63Hz	104.0	103.179	102.3	102.3	dB(A)	-0.85	-0.85
125Hz	104.0	103.828	103.1	103.1	dB(A)	-0.70	-0.70
250Hz	104.0	103.999	103.3	103.3	dB(A)	-0.67	-0.67
500Hz	104.0	104.033	103.5	103.5	dB(A)	-0.51	-0.51
1kHz	104.0	104.000	103.3	103.5	dB(A)	-0.67	-0.50
2kHz	104.0	103.830	103.0	103.5	dB(A)	-0.80	-0.36
4kHz	104.0	103.175	101.7	102.8	dB(A)	-1.43	-0.36
8kHz	104.0	100.955	97.8	101.1	dB(A)	-3.13	0.11
12.5kHz	104.0	97.828	91.2	97.6	dB(A)	-6.78	-0.21
16kHz	104.0	95.371	87.2	95.2	dB(A)	-8.57	-0.16





# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER

**160922**

## RESULT SHEET 0265 - 820 SOUND LEVEL METER

### 2) ACOUSTIC LEVEL

Lzf

Applied Freq	Applied Level	Limits	Indicated	Units	Error %
1kHz	94.0	±2%	93.5	dB	-0.53
1kHz	104.0	±2%	103.3	dB	-0.67
1kHz	114.0	±2%	113.5	dB	-0.44

# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER

**160922**


## RESULT SHEET 0265 - 820 SOUND LEVEL METER

### POST ADJUSTMENT

#### 1) FREQUENCY RESPONSE

##### A Weighted

Applied Freq	Applied Level	A Weighting Typical	Indicated Pressure Response	Free Field Response	Units	Error % Pressure	Error % FFR
31.5Hz	104.0	64.471	64.5	64.5	dBA	0.04	0.04
63Hz	104.0	77.777	77.8	77.8	dBA	0.03	0.03
125Hz	104.0	87.811	88.1	88.1	dBA	0.33	0.33
250Hz	104.0	95.325	95.7	95.7	dBA	0.39	0.39
500Hz	104.0	100.752	101.1	101.1	dBA	0.35	0.35
1kHz	104.0	104.000	104.0	104.2	dBA	0.00	0.17
2kHz	104.0	105.202	105.0	105.5	dBA	-0.19	0.25
4kHz	104.0	104.964	104.1	105.2	dBA	-0.82	0.22
8kHz	104.0	102.855	100.3	103.6	dBA	-2.48	0.70
12.5kHz	104.0	99.750	93.7	100.1	dBA	-6.07	0.37
16kHz	104.0	97.300	89.7	97.7	dBA	-7.81	0.43



# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910




CERTIFICATE NUMBER

**160922**

## RESULT SHEET 0265 - 820 SOUND LEVEL METER

### 1) FREQUENCY RESPONSE (Continued) C Weighted

Applied Freq	Applied Level	C Weighting Typical	Indicated Pressure Response	Free Field Response	Units	Error % Pressure	Error % FFR
31.5Hz	104.0	100.969	100.5	100.5	dBA	-0.46	-0.46
63Hz	104.0	103.179	103.0	103.0	dBA	-0.17	-0.17
125Hz	104.0	103.828	103.7	103.7	dBA	-0.12	-0.12
250Hz	104.0	103.999	103.9	103.9	dBA	-0.10	-0.10
500Hz	104.0	104.033	104.0	104.0	dBA	-0.03	-0.03
1kHz	104.0	104.000	104.0	104.2	dBA	0.00	0.17
2kHz	104.0	103.830	103.5	104.0	dBA	-0.32	0.13
4kHz	104.0	103.175	102.3	103.4	dBA	-0.85	0.22
8kHz	104.0	100.955	98.4	101.7	dBA	-2.53	0.71
12.5kHz	104.0	97.828	91.8	98.2	dBA	-6.16	0.40
16kHz	104.0	95.371	87.8	95.8	dBA	-7.94	0.47





# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER

**160922**

## RESULT SHEET 0265 - 820 SOUND LEVEL METER

### 2) ACOUSTIC LEVEL

Lzf

Applied Freq	Applied Level	Limits	Indicated	Units	Error %
1kHz	94.0	±2%	94.0	dB	0.00
1kHz	104.0	±2%	104.0	dB	0.00
1kHz	114.0	±2%	114.0	dB	0.00

BATTERIES REPLACED

NO

	Model	Serial Number
Sound Level Meter	820	1350
Pre-Amplifier	PRM828	1568
Microphone	40AE	37024

### COMMENTS

Calibration carried out in accordance with BS7580 part 1 1997, section 5.6.1.

TEST ENGINEER **M A FROST**

DATE **02 JUNE 2016**

**Calibration Certificates for Position N2 – August 2016 Survey**





# CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER

**162034**

**Page 2 of 5**

## INSTRUMENTS USED

<b>EQUIPMENT</b>	<b>SERIAL No</b>	<b>CERTIFICATE No</b>	<b>CAL DUE</b>
Bruel & Kjaer 4226	2952856	U19912	26 Oct 2016

## Notes:

## MEASUREMENT UNCERTAINTIES

The expanded uncertainty quoted refers to the measured values only, with no account being taken of the instruments ability to maintain its calibration. The expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%.

<b>PARAMETER</b>	<b>RANGE</b>	<b>EXPANDED UNCERTAINTY</b>
d.c. Resistance	0.01 $\Omega$ - 400M $\Omega$	$\pm$ 409ppm
	400M $\Omega$ - 1T $\Omega$	$\pm$ 1%
d.c. Voltage	0V - 1kV	$\pm$ 79ppm
d.c. Voltage	1.01kV-15kV	$\pm$ 2.2%
d.c. Current	0mA - 20A	$\pm$ 437ppm
a.c. Voltage	0mV - 1.05kV	$\pm$ 1.2%
a.c. Current	0mA - 20A	$\pm$ 0.5%
Frequency	0.5Hz - 20GHz	$\pm$ 0.1ppm
Capacitance	0.5nF - 40mF	$\pm$ 1.1%
Time	0 - 1 Hour	$\pm$ 1s
Distortion	10mV - 100V	$\pm$ 1.4mV
Temperature (Dry Block)	-30 $^{\circ}$ C - 350 $^{\circ}$ C	$\pm$ 1%
Temperature (Simulation)	-270 $^{\circ}$ C - 1800 $^{\circ}$ C	$\pm$ 0.57%
Pressure	10mBar - 35Bar	$\pm$ 0.04%
Torque	0.1 - 1100Nm	$\pm$ 0.5%
Weight	2g - 157kg	$\pm$ 0.03%
Humidity	0% - 90%	$\pm$ 1%
Shock & Impulse Hammers		$\pm$ 4%
Spring Hammers		$\pm$ 0.015J
Sound	Frequency	$\pm$ 0.06%
Sound	Level	$\pm$ 0.16dB
Tachometers	60rpm - 96000rpm	$\pm$ 0.1%
Anemometers	2.5m/s to 15m/s	$\pm$ 2.0%
Vibration Meters	10Hz - 1kHz	$\pm$ 5%
Vibration Calibrators		$\pm$ 3%
Mechanical Measurement	<200mm	$\pm$ 0.0003mm
	>200mm	$\pm$ 0.002mm
Inductance		$\pm$ 0.1%
Power (VA)		$\pm$ 1%
Power (RF)		$\pm$ 0.5dB

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER

**162034**


## RESULT SHEET 0265 - 820 SOUND LEVEL METER

### AS FOUND

### 1) FREQUENCY RESPONSE - **UPDATE FFR VALUES - Currently GRAS 40AE**

#### A Weighted

Applied Freq	Applied Level	A Weighting Typical	Indicated Pressure Response	Free Field Response	Units	Error % Pressure	Error % FFR
31.5Hz	104.0	64.471	64.3	64.3	dBA	-0.27	-0.27
63Hz	104.0	77.777	77.5	77.5	dBA	-0.36	-0.36
125Hz	104.0	87.811	87.6	87.6	dBA	-0.24	-0.24
250Hz	104.0	95.325	95.0	95.0	dBA	-0.34	-0.34
500Hz	104.0	100.752	100.4	100.4	dBA	-0.35	-0.35
1kHz	104.0	104.000	103.6	103.8	dBA	-0.38	-0.21
2kHz	104.0	105.202	104.4	104.9	dBA	-0.76	-0.33
4kHz	104.0	104.964	103.5	104.6	dBA	-1.39	-0.35
8kHz	104.0	102.855	100.0	103.3	dBA	-2.78	0.40
12.5kHz	104.0	99.750	95.6	102.0	dBA	-4.16	2.28
16kHz	104.0	97.300	91.7	99.7	dBA	-5.76	2.49



# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER


**162034**

## RESULT SHEET 0265 - 820 SOUND LEVEL METER

### 1) FREQUENCY RESPONSE (Continued)

#### C Weighted

Applied Freq	Applied Level	C Weighting Typical	Indicated Pressure Response	Free Field Response	Units	Error % Pressure	Error % FFR
31.5Hz	104.0	100.969	100.9	100.9	dBA	-0.07	-0.07
63Hz	104.0	103.179	103.1	103.1	dBA	-0.08	-0.08
125Hz	104.0	103.828	103.8	103.8	dBA	-0.03	-0.03
250Hz	104.0	103.999	103.9	103.9	dBA	-0.10	-0.10
500Hz	104.0	104.033	104.0	104.0	dBA	-0.03	-0.03
1kHz	104.0	104.000	103.7	103.9	dBA	-0.29	-0.12
2kHz	104.0	103.830	103.6	104.1	dBA	-0.22	0.22
4kHz	104.0	103.175	103.5	104.6	dBA	0.31	1.38
8kHz	104.0	100.955	99.7	103.0	dBA	-1.24	2.00
12.5kHz	104.0	97.828	93.7	100.1	dBA	-4.22	2.34
16kHz	104.0	95.371	89.9	97.9	dBA	-5.74	2.67



# CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER

**162034**

## RESULT SHEET 0265 - 820 SOUND LEVEL METER

### 2) ACOUSTIC LEVEL

Lzf

Applied Freq	Applied Level	Limits	Indicated	Units	Error %
1kHz	94.0	±2%	94.0	dB	0.00
1kHz	104.0	±2%	104.0	dB	0.00
1kHz	114.0	±2%	114.1	dB	0.09

**BATTERIES REPLACED**

**YES**

	Model	Serial Number
Sound Level Meter	824	824A1419
Pre-Amplifier	PRM902	2732
Microphone	40AE	31817

### COMMENTS

Calibration carried out in accordance with BS7580 part 1 1997, section 5.6.1.

**TEST ENGINEER R J WADE**

**DATE 04 AUGUST 2016**



# FAULT REPORT

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

BS EN ISO 9001:2008 APPROVAL CERTIFICATE No. 953910



CERTIFICATE NUMBER

**162034**

**Page 1 of 1**

<b><u>Customer</u></b>	SPL ACOUSTICS LTD
<b><u>Order No</u></b>	EQP/CAL/2016-0802
<b><u>Equipment Description</u></b>	SOUND LEVEL METER
<b><u>Manufacturer</u></b>	LARSON DAVIS
<b><u>Model</u></b>	824
<b><u>Serial No</u></b>	824A1419
<b><u>Ident No</u></b>	SPL A

## **DETAILS**

Battery terminals were so corroded that the unit would not switch on.  
Terminals were cleaned as best as possible.  
Unit now functions correctly.

**TEST ENGINEER R J WADE**

**DATE 4 AUGUST 2016**

**Calibration Certificates for Position N3 – August 2016 Survey**

# Chapitre 1.

## CONSTAT DE VERIFICATION

### VERIFICATION CERTIFICATE

---

CV-DTE-L-15-PVE-32321

DELIVRE PAR :  
ISSUED BY :

ACOEM  
Service Métrologie

69760 LIMONEST  
France

INSTRUMENT VERIFIE  
INSTRUMENT CHECKED

Désignation :  
Designation :

**Sonomètre Intégrateur-Moyenneur**  
**Integrating-Averaging Sound Level Meter**

Constructeur :  
Manufacturer :

**01dB**

Type :  
Type :

**CUBE**

N° de serie :  
Serial number :

**10692**


N° d'identification :  
Identification number

Date d'émission :  
Date of issue :

**18/02/15**

Ce constat comprend 5 pages  
This certificate includes pages

LE RESPONSABLE METROLOGIQUE  
DU LABORATOIRE  
HEAD OF THE METROLOGY LAB



Jean-Claude Périchon

LA REPRODUCTION DE CE CONSTAT N'EST AUTORISEE  
QUE SOUS LA FORME DE FAC-SIMILE PHOTOGRAPHIQUE INTEGRAL

CE DOCUMENT NE PEUT PAS ETRE UTILISE EN LIEU  
ET PLACE D'UN CERTIFICAT D'ETALONNAGE. CE DOCUMENT  
EST REALISE SUIVANT LES RECOMMANDATIONS DU  
FASCICULE DE DOCUMENTATION X 07-011.

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED OTHER  
THAN IN FULL BY PHOTOGRAPHIC PROCESS

THIS DOCUMENT CAN'T BE USED AS CALIBRATION  
CERTIFICATE. IT IS COMPLIANT WITH THE X 07-011 STANDARD  
RECOMMENDATIONS.

**Calibration Certificates for Position N5 – August 2016 Survey**

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801



0801

University of  
**Salford**  
MANCHESTER

Page 1 of 2

## APPROVED SIGNATORIES

Claire Lomax [x]      Andy Moorhouse [ ]

Gary Phillips [ ]      Danny McCaul [ ]

## acoustic calibration laboratory

The University of Salford, Salford, Greater Manchester, M5 4WT, UK  
<http://www.acoustics.salford.ac.uk>  
t 0161 295 3030/0161 295 3319 f 0161 295 4456 e c.lomax1@salford.ac.uk

Certificate Number: 02678/2

Date of Issue: 28 April 2016

## VERIFICATION OF A TYPE 1 SOUND LEVEL METER to BS7580 Part 1

FOR:	Resound Acoustics Limited 13 Rother Street Stratford-upon-Avon Warwickshire CV37 6LU
FOR THE ATTENTION OF:	Mike Brownstone
CALIBRATION DATE:	27 <sup>th</sup> and 28 <sup>th</sup> April 2016
TEST PROCEDURE:	CTP08 (Laboratory Manual)

<b>Sound Level Meter</b>			
Manu:	01dB	Model:	Solo      Serial No: 65682
<b>Microphone</b>			
Manu:	01dB	Model:	MCE212      Serial No: 153491
<b>Preamp</b>			
Manu:	01dB	Model:	PRE 21 S      Serial No: 16310
<b>Associated Calibrator</b>			
Manu:	01 dB	Model:	CAL 21      Serial No: 34134139 (2013)      Adaptor: BAC21

Test Engineer (initial): GP

Name: Gary Phillips



# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 2 of 2

Certificate Number: 02678/2

Date of Issue: 28 April 2016

## SET-UP INFORMATION

The instrument version was Master 01 V1.405. The reference range, reference SPL, primary indicator range, pulse range and linearity range as specified by the manufacturer have been used. The instrument was adjusted to read 93.9 dB (A) in response to the associated calibrator. This reading was obtained from the calibration certificate of the calibrator, 02678/1 and information in the manufacturer's instruction manual. The instrument was calibrated without a windshield. Consult manufacturer's instructions if using a windshield.

## MEASUREMENTS

The levels of self-generated noise were:

<b>A:</b>	<b>11.7 dB*</b>
<b>B:</b>	<b>10.9 dB*</b>
<b>C:</b>	<b>13.4 dB*</b>
<b>Z:</b>	<b>17.2 dB*</b>

\*Under-range indicated on instrument display.

At the end of the tests the indication of the sound level meter in response to the associated sound calibrator was 94.0 dB (A) which corresponds to the following level at 101.325 kPa:

**Sound Pressure Level                      94.0 dB (A)**

**This reading should be used henceforth to set up the sound level meter for field use.**

THE SOUND LEVEL METER WAS VERIFIED ACCORDING TO THE PROCEDURE GIVEN IN BS7580: Part 1 1997 WITH THE FOLLOWING EXCEPTIONS:

The microphone corrections applied as specified in BS 7580: Part 1: 1997 were obtained from a frequency response measurement by this Laboratory using the electrostatic actuator method. The response in isolation is not covered by our UKAS accreditation.

A stricter test than that specified in 5.5.10 and 5.5.11 of BS 7580 has been used by not applying the low level signal.

## STATEMENT OF RESULT:

THE SOUND LEVEL METER CONFORMS TO THE TYPE 1 REQUIREMENTS OF BS7580: PART1 1997.

Instruments used in the verification procedure were traceable to National Standards. The method of acoustic calibration employed a standard sound pressure calibrator for the 1 kHz test whilst the tests at 125 Hz and 8 kHz were performed by the electrostatic actuator method. The uncertainty of the Laboratory's 1 kHz calibrator was  $\pm 0.10$  dB. The uncertainty of the standard calibrator is not included in the applied tolerances. It is assumed that the sound level meter was manufactured in accordance with BSEN60651: 1994 Type 1, and BSEN60804: 1994 Type 1.

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements. All measurement results are retained at the acoustic calibration laboratory for at least four years.*

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 1 of 3

## APPROVED SIGNATORIES

Claire Lomax [ ]      Andy Moorhouse [ ]

Gary Phillips [x]      Danny McCaul [ ]

## acoustic calibration laboratory

The University of Salford, Salford, Greater Manchester, M5 4WT, UK  
<http://www.acoustics.salford.ac.uk>  
t 0161 295 3030/0161 295 3319 f 0161 295 4456 e c.lomax1@salford.ac.uk



UKAS  
CALIBRATION

0801

University of  
**Salford**  
MANCHESTER

Certificate Number: 02678/3

Date of Issue: 28 April 2016

## VERIFICATION OF A SOUND LEVEL METER / ANALYSER 1/3 OCTAVE FILTER SET MANUFACTURED TO BS EN 61260: 1996 RELATIVE ATTENUATION

FOR: Resound Acoustics Limited  
13 Rother Street  
Stratford-upon-Avon  
Warwickshire  
CV37 6LU

FOR ATTENTION OF: Mike Brownstone

DESCRIPTION: Sound level meter/analyser version, Master 01 V1.405 with  
1/3 octave filter set manufactured to BS EN 61260: 1996.

DATE OF CALIBRATION: 27/04/2016

TEST PROCEDURE: CTP19 (Laboratory Manual)

### Sound Level Meter/Analyser details

Manu: 01dB      Model: Solo      Serial No: 65682

### Filter details

Filter Base: 2      Filter Class: 1

Test Engineer (initial): CL

Name: Claire Lomax

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 2 of 3

Certificate Number: 02678/3

Date of Issue: 28 April 2016

## INSTRUMENT SET UP

The instrument was adjusted to read 93.9 dB (A) in response to the associated calibrator. This reading was obtained from the calibration certificate of the calibrator, 02678/1 and information in the manufacturer's instruction manual. The instrument was calibrated without a windshield. Consult manufacturer's instructions if using a windshield. The instrument was set to measure SPL on the Z-weighting in each of the tested frequency bands. Exact base 2 frequencies have been applied in all of the tests. All tests were performed on the reference level range. The manufacturer's instruction manual available did not specify the upper limit of the Linear Operating range, so an upper limit of 130 dB has been assumed and signals applied at a level 1 dB below the assumed upper limit. The test signals were applied to the preamplifier of the instrument, a PRE 21 S, serial number 16310, via an appropriate input adaptor.

## RELATIVE ATTENUATION TESTS

**The Relative Attenuation of the combination of filter set and sound level meter was tested over the following frequency ranges:-**

1/3 octave filter bands from 20 Hz to 20 kHz

## RELATIVE ATTENUATION TEST RESULTS

Class 1 tolerances: Table 1 of BS EN 61260:1996

**Results: 20 Hz to 12.5 kHz Pass.**

**NOTE: The 16 kHz and 20 kHz filter bands both failed the Relative Attenuation test at a single test point. See results on page 3 of this certificate for the actual results. However, this failure is in line with the manufacturer expectations given in Chapter 2, Page 12 of the manufacturer instruction manual, reference: DOC1110 December 2012 C - SOLO Black Edition Technical Manual GB.**

Uncertainty of measurement within filter pass-band: 0.19 dB coverage factor  $k=2$

Uncertainty of measurement outside filter pass-band: 0.21 dB coverage factor  $k=2$

### NOTE:

These results apply only to the tested filter bands and do not imply that any untested filter bands would also pass the reported test. The results are only valid for the combination of filter set and sound level meter / analyser tested.

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.*

*All measurement results are retained at the acoustic calibration laboratory for at least four years.*

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*

# Certificate of Calibration

Issued by University of Salford (Acoustics Calibration Laboratory)  
UKAS ACCREDITED CALIBRATION LABORATORY NO. 0801

Page 3 of 3

Certificate Number: 02678/3

Date of Issue: 28 April 2016

## 16 kHz and 20 kHz Relative Attenuation test results.

Test 1/3 octave filter X= 12 fexact=16000.0 Hz class 1

Nominal	Measured	LoLim	HiLim	Result
f[Hz]	L[dB]	[dB]	[dB]	[Pass/Fail]
2944.02	41.1	0	59	Pass
5212.5	55.3	0	68	Pass
8479.3	75.8	0	87	Pass
12349	104.2	0	111.5	Pass
14254.4	125.4	124	127	Pass
14709.1	128.2	127.7	129.3	Pass
15152.4	128.8	128.4	129.3	Pass
15583	128.8	128.6	129.3	Pass
16000	128.9	128.7	129.3	Pass
16428.1	128.9	128.6	129.3	Pass
16895	128.9	128.4	129.3	Pass
17404.2	128.8	127.7	129.3	Pass
17959.4	126.1	124	127	Pass
20730.4	88.2	0	111.5	Pass
30191.2	0	0	87	Pass
49112.7	3.8	0	68	Pass
<b>86955.9</b>	<b>64.9</b>	<b>0</b>	<b>59</b>	<b>FAIL</b>

Test 1/3 octave filter X= 13 fexact=20158.7 Hz class 1

Nominal	Measured	LoLim	HiLim	Result
f[Hz]	L[dB]	[dB]	[dB]	[Pass/Fail]
3709.24	45.4	0	59	Pass
6567.33	60.9	0	68	Pass
10683.2	82.2	0	87	Pass
15558.8	108.4	0	111.5	Pass
17959.4	125.8	124	127	Pass
18532.3	128.2	127.7	129.3	Pass
19090.8	128.9	128.4	129.3	Pass
19633.4	129	128.6	129.3	Pass
20158.7	129.1	128.7	129.3	Pass
20698.2	129.1	128.6	129.3	Pass
21286.4	129	128.4	129.3	Pass
21927.9	129	127.7	129.3	Pass
22627.4	126.2	124	127	Pass
26118.7	11.6	0	111.5	Pass
38038.5	11	0	87	Pass
61878.2	19.9	0	68	Pass
<b>109558</b>	<b>61.8</b>	<b>0</b>	<b>59</b>	<b>FAIL</b>

*This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.*